

# CTS-21240

Acquity® I-Class SM-FTN Performance Maintenance Kit

\*Injector Valve Cartridge (289006464) not included

\*Rotor Seal & Pod Repair Kit included

Comparable to: Waters® 201000259



## Technical Data Sheet

### 1.0 Description

This kit contains all the parts necessary to keep Waters Acquity® I-Class SM-FTN systems working at peak performance. It is comparable to Waters® part number 201000259.

### 2.0 Specifications

#### 2.1 Components

Qty	Part Description
1	Rotor Seal & Pod Repair Kit, Vespel®, 3-Groove for ACQUITY Everflow Injector Pod/Cartridge 289006464 (11-4134)
1	Assy, Needle, FTN-15 uL, SST, 26L (CTS-21078)
1	Needle Guide (11-2591)
1	Needle Seat (11-2570)
1	Assembly, Seat Port, 0.003 ID (CTS-21073)
1	Assy, Tubing, BSM Outlet to SM (CTS-21075)
2	Ferrule, 1/16", SS (CTS-2330)
1	1/16" Ferrule Set, 2/pk (CTS-21243)
1	100µL Sample Metering Syringe, HP (CTS-A11167)
1	Filter, Air, Side Panel (11-3606)



#### 2.2 Wetted Materials

PTFE  
316 SS  
HF1 Foam  
PEEK™  
Vespel®  
Glass  
Neoprene

### 3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.

### 4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.