

# CTS-10573

Injection Valve Rotor Seal  
Comparable to: Hitachi®  
ANO-0818



## Technical Data Sheet

---

### 1.0 Description


This rotor seal is designed to be compatible with Hitachi AS-7200 and AS-7250 HPLC systems. It is comparable to Hitachi part number ANO-0818.

### 2.0 Materials Specifications

304L SS  
UHMWPE  
Vespel®

### 3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.