

CTS-3109

250µL Syringe
Comparable to: Waters®
WAT073109



Technical Data Sheet

1.0 Description

This syringe is designed to be compatible with Waters 717, 717 Plus, 2690, 2690D, 2695, 2695D, 2790, 2795 and Alliance HPLC systems. It is comparable to Waters part number WAT073109.

2.0 Materials Specifications

316 SS
Borosilicate Glass
PTFE

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.