

CTS-21802

Sample Loop, 100 μ L, SS
Comparable to Agilent® G7129-87302



Technical Data Sheet

1.0 Description


This sample loop is designed to be compatible with Agilent® 1260 and 1290 HPLC systems. It is comparable to Agilent part number G7129-87302.

2.0 Materials Specifications

PTFE Color Code Tubing
316 SS

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.