

CTS-10686

Needle Seal, Vespel®
Comparable to: Shimadzu®
228-33355-04



Technical Data Sheet

1.0 Description

This needle seal is designed to be compatible with Shimadzu® LC-2010, SIL-10ADvp and SIL-10AXL HPLC systems. It is comparable to Shimadzu® part number 228-33355-04.

2.0 Wetted Materials Specifications

Vespel®

3.0 Replacement Frequency

Need for replacement of the needle seal may be made on a preventive maintenance schedule or when it fails to perform to specification.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.