

11-4148

PM Kit, G1329B
Comparable to: Agilent®
G1313-68719



Technical Data Sheet

1.0 Description

This kit contains all the parts necessary to keep Agilent 1100, 1120, 1200, 1220, 1260, G1313A, and G1329A/B HPLC systems working at peak performance. It is comparable to Agilent part number G1313-68719.

2.0 Specifications

2.1 Components

Qty	Part Description
1	Seal, Pump (CTS-10621)
1	Needle Assembly (CTS-10835)
1	2 Position/6 Port Rotor Seal, Vespel® (600 bar) (CTS-12012)
1	Needle Seat, Std., PEEK™, 0.17mm ID (11-2605)

2.2 Materials

316 SS
316 SS Spring
316Ti SS
Carbon Filled PTFE
PEEK
Vespel

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.