

CTS-22028

PEEK™ Adapter, Female, 1/4-28 to 10-32,
Comparable to: Agilent® 0100-1847



Technical Data Sheet

1.0 Description

This adapter is designed to be compatible with Agilent® 1100, 1200, and 1260 HPLC systems. It is comparable to Agilent® part number 0100-1847.

2.0 Materials Specifications

PEEK™

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.

Disclaimer: Sciencix, Inc. is not affiliated with or an authorized distributor/official partner of Original Equipment Manufacturer (OEM) parts. All OEM part names, numbers, brands, registered trademarks, or trademarks are the sole property of their owners in the US or any other country. All OEM part names, numbers, and company names used in this report are for cross reference purposes only. Use of these names and brands does not imply endorsement. Statements made by Sciencix regarding the quality and performance of our products are our own and should not be construed as an indication of OEM acceptance or approval.