

CTS-21782

Piston Seal, 900 µL
Comparable to: Agilent® 0905-1294



Technical Data Sheet

1.0 Description

This Piston Seal is designed to be compatible with Agilent® 1100, G1313A, G1329A/B, G1367E instruments. It is comparable to Agilent® part number 0905-1294.

2.0 Materials Specifications

PTFE

316 Stainless Steel

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.