

CTS-3209

Syringe Seal, 250 μ L/2500 μ L
Comparable to: Waters®
WAT077347



Technical Data Sheet

1.0 Description


This syringe seal is designed to be compatible with Waters 717, 717 Plus, 2690, 2690D, 2695, 2695D, 2790, 2795 and Alliance HPLC systems. It is comparable to Waters part number WAT077347.

2.0 Materials Specifications

PTFE

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.