

CTS-21422

Needle Seal Kit

Comparable to: Shimadzu® 228-52401-92



Technical Data Sheet

1.0 Description

This kit contains all the parts necessary to keep Shimadzu® SIL-30AC HPLC systems working at peak performance. It is comparable to Shimadzu® part number 228-52401-92.

2.0 Specifications

2.1 Components

Qty	Part Description
1	Seal Sleeve (CTS-21088)
3	Needle Seal (11-4052)



2.2 Materials

Polyimide
316 SST

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.

4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.