

CTS-10391

Plunger Wash Seal
Comparable to: Beckman® 728772,
238627



Technical Data Sheet

1.0 Description

This seal is designed to be compatible with Beckman 125, 126, 127 and 128 HPLC systems. It is comparable to Beckman part numbers 728772 and 238627.

2.0 Materials Specifications

Carbon Filled PTFE
300 SS

3.0 Replacement Frequency

Beyond routine maintenance it may be necessary to replace the seals more frequently if they are operating in a high salt buffer concentration environment; or if the plungers are scratched or worn. Seals may also need replacing if the system fails its leak test.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.