

CTS-21277

Needle Seat, 0.12, ID, 0.8mm OD, PEEK, 600 Bar
Comparable to: Agilent® G1367-87012



Technical Data Sheet

1.0 Description

This needle seat is designed to be compatible with 1260 systems. It is comparable to Agilent® part number G1367-87012.

2.0 Materials Specifications

PEEK™

Buna-N

316 SS

PTFE

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.